

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
 Data File : VX023515.D
 Acq On : 29 Jul 2021 15:58
 Operator : JC/MD
 Sample : M3216-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPLP-LEACHATE-AMENDED-COM-

Quant Time: Jul 30 03:43:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	165313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	280692	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102174	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107302	52.034	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.060%	
35) Dibromofluoromethane	5.397	113	74351	42.365	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 84.720%	
50) Toluene-d8	8.653	98	336939	52.255	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.520%	
62) 4-Bromofluorobenzene	11.085	95	113361	51.702	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.400%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	15364	15.188	ug/l	# 87
20) Methylene Chloride	2.794	84	9652	4.905	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073021\
Data File : VX023515.D
Acq On : 29 Jul 2021 15:58
Operator : JC/MD
Sample : M3216-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPLP-LEACHATE-AMENDED-COM-

Quant Time: Jul 30 03:43:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

