

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073025\
 Data File : VX047185.D
 Acq On : 30 Jul 2025 14:20
 Operator : JC/MD
 Sample : Q2709-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 7211935-295

Quant Time: Jul 31 03:16:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	330471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	562738	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	506567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	255879	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	172705	41.953	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.900%
35) Dibromofluoromethane	5.403	113	146409	42.138	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.280%
50) Toluene-d8	8.653	98	450490	40.036	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	80.080%#
62) 4-Bromofluorobenzene	11.079	95	208287	42.954	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.900%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	46094	27.951	ug/l	98
18) Methyl Acetate	2.715	43	30597	6.982	ug/l	97
20) Methylene Chloride	2.806	84	8658	1.908	ug/l	94
37) Ethyl Acetate	4.739	43	17740	3.251	ug/l	98
43) Isopropyl Acetate	6.367	43	13906	1.634	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073025\
Data File : VX047185.D
Acq On : 30 Jul 2025 14:20
Operator : JC/MD
Sample : Q2709-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
7211935-295

Quant Time: Jul 31 03:16:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

