

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073025\
 Data File : VX047186.D
 Acq On : 30 Jul 2025 15:00
 Operator : JC/MD
 Sample : Q2709-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 7211935-295

Quant Time: Jul 31 03:17:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	302963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	516099	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	457505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	230741	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	204834	54.276	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.560%	
35) Dibromofluoromethane	5.391	113	173716	54.515	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.040%	
50) Toluene-d8	8.647	98	529897	51.348	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.700%	
62) 4-Bromofluorobenzene	11.079	95	240626	54.108	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.220%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	44844	29.662	ug/l	99
18) Methyl Acetate	2.715	43	31869	7.932	ug/l	99
20) Methylene Chloride	2.800	84	8945	2.151	ug/l	90
37) Ethyl Acetate	4.733	43	17972	3.591	ug/l #	94
43) Isopropyl Acetate	6.348	43	15024	1.925	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073025\
Data File : VX047186.D
Acq On : 30 Jul 2025 15:00
Operator : JC/MD
Sample : Q2709-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
7211935-295

Quant Time: Jul 31 03:17:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

