

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073119\
 Data File : VX011253.D
 Acq On : 31 Jul 2019 19:13
 Operator : JC/SP
 Sample : K4048-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWL-C2-GW

Quant Time: Aug 01 03:27:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	233620	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	336029	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	298446	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137974	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118730	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
35) Dibromofluoromethane	5.50	113	99269	46.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.08%	
50) Toluene-d8	8.71	98	391488	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
62) 4-Bromofluorobenzene	11.13	95	130188	44.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.56%	
Target Compounds						
73) Isopropylbenzene	11.02	105	3036	0.326	ug/l	90
83) tert-Butylbenzene	11.77	119	2544	0.332	ug/l	84
85) sec-Butylbenzene	11.94	105	3453	0.385	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073119\
Data File : VX011253.D
Acq On : 31 Jul 2019 19:13
Operator : JC/SP
Sample : K4048-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWL-C2-GW

Quant Time: Aug 01 03:27:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

