

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073120\
 Data File : VX017705.D
 Acq On : 31 Jul 2020 16:38
 Operator : JC/SP
 Sample : L3526-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB179-GW-198-200

Quant Time: Aug 01 03:59:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	478510	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	821707	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	878036	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	481542	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	331869	53.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.98%	
35) Dibromofluoromethane	5.47	113	276817	50.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.74%	
50) Toluene-d8	8.70	98	1011875	51.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.08%	
62) 4-Bromofluorobenzene	11.12	95	454173	57.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.48%	
Target Compounds						
16) Acetone	2.42	43	18514	6.843	ug/l	90
27) cis-1,2-Dichloroethene	4.57	96	281637	48.972	ug/l	88
44) Trichloroethene	7.19	130	948341	143.623	ug/l	98
52) Toluene	8.76	92	4424	0.305	ug/l #	73
64) Tetrachloroethene	9.32	164	118648	15.056	ug/l	98
68) m/p-Xylenes	10.34	106	3278	0.267	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073120\
Data File : VX017705.D
Acq On : 31 Jul 2020 16:38
Operator : JC/SP
Sample : L3526-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB179-GW-198-200

Quant Time: Aug 01 03:59:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

