

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073120\
 Data File : VX017706.D
 Acq On : 31 Jul 2020 17:01
 Operator : JC/SP
 Sample : L3526-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB179-GW-218-220

Quant Time: Aug 01 04:02:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	461441	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	754942	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	816912	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	478007	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	320772	54.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.24%	
35) Dibromofluoromethane	5.47	113	266255	52.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.46%	
50) Toluene-d8	8.70	98	949519	52.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.26%	
62) 4-Bromofluorobenzene	11.12	95	456866	62.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.34%	
Target Compounds						
16) Acetone	2.42	43	15628	5.990	ug/l #	88
27) cis-1,2-Dichloroethene	4.57	96	131350	23.684	ug/l	86
44) Trichloroethene	7.19	130	508301	83.789	ug/l	100
64) Tetrachloroethene	9.32	164	117598	16.039	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073120\
Data File : VX017706.D
Acq On : 31 Jul 2020 17:01
Operator : JC/SP
Sample : L3526-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB179-GW-218-220

Quant Time: Aug 01 04:02:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

