

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073120\
 Data File : VX017709.D
 Acq On : 31 Jul 2020 18:09
 Operator : JC/SP
 Sample : L3526-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB179-GW-278-280

Quant Time: Aug 01 04:13:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	467926	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	778837	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	838858	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	460066	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	323428	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
35) Dibromofluoromethane	5.47	113	273739	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
50) Toluene-d8	8.70	98	970722	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
62) 4-Bromofluorobenzene	11.12	95	441661	58.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.46%	
Target Compounds						
12) 1,1-Dichloroethene	2.37	96	5488	1.135	ug/l #	85
16) Acetone	2.42	43	17577	6.644	ug/l	89
19) Methyl tert-butyl Ether	3.17	73	26869	1.606	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	65759	11.693	ug/l	85
32) 1,1,1-Trichloroethane	5.46	97	6617	0.739	ug/l #	49
44) Trichloroethene	7.19	130	406444	64.943	ug/l	94
64) Tetrachloroethene	9.32	164	34808	4.623	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073120\
Data File : VX017709.D
Acq On : 31 Jul 2020 18:09
Operator : JC/SP
Sample : L3526-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB179-GW-278-280

Quant Time: Aug 01 04:13:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

