

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073120\
 Data File : VX017729.D
 Acq On : 01 Aug 2020 02:53
 Operator : JC/SP
 Sample : PB130661ZHE#25
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130661ZHE#25

Quant Time: Aug 01 05:21:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	448805	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	758530	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	790728	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	429532	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	321150	55.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.42%	
35) Dibromofluoromethane	5.47	113	267967	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
50) Toluene-d8	8.70	98	941561	51.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.90%	
62) 4-Bromofluorobenzene	11.12	95	405847	55.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.82%	
Target Compounds						
16) Acetone	2.42	43	13715	5.405	ug/l	99
20) Methylene Chloride	2.84	84	19513	2.873	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073120\
Data File : VX017729.D
Acq On : 01 Aug 2020 02:53
Operator : JC/SP
Sample : PB130661ZHE#25
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130661ZHE#25

Quant Time: Aug 01 05:21:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

