

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073120\
 Data File : VX017731.D
 Acq On : 01 Aug 2020 03:39
 Operator : JC/SP
 Sample : PB130661ZHE#27
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130661ZHE#27

Quant Time: Aug 01 05:25:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	434191	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	720526	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	786934	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	421172	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	309974	55.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.16%	
35) Dibromofluoromethane	5.46	113	259197	53.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.58%	
50) Toluene-d8	8.70	98	894233	51.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.88%	
62) 4-Bromofluorobenzene	11.12	95	398043	57.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.42%	
Target Compounds						
16) Acetone	2.42	43	13892	5.659	ug/l	89
20) Methylene Chloride	2.84	84	19536	3.019	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073120\
 Data File : VX017731.D
 Acq On : 01 Aug 2020 03:39
 Operator : JC/SP
 Sample : PB130661ZHE#27
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 PB130661ZHE#27

Quant Time: Aug 01 05:25:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

