

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073120\
 Data File : VX017732.D
 Acq On : 01 Aug 2020 04:02
 Operator : JC/SP
 Sample : PB130661ZHE#28
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130661ZHE#28

Quant Time: Aug 01 05:27:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	435593	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	724205	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	787863	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	467214	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	307054	54.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.76%	
35) Dibromofluoromethane	5.47	113	265399	54.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.58%	
50) Toluene-d8	8.70	98	899223	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	11.12	95	417015	59.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.28%	
Target Compounds						
16) Acetone	2.42	43	12810	5.202	ug/l #	85
20) Methylene Chloride	2.84	84	17091	2.464	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073120\
Data File : VX017732.D
Acq On : 01 Aug 2020 04:02
Operator : JC/SP
Sample : PB130661ZHE#28
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130661ZHE#28

Quant Time: Aug 01 05:27:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

