

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX073120\
 Data File : VX017733.D
 Acq On : 01 Aug 2020 04:25
 Operator : JC/SP
 Sample : PB130661ZHE#29
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130661ZHE#29

Quant Time: Aug 01 05:29:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	436927	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	758301	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	791771	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	433936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	312271	55.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.28%	
35) Dibromofluoromethane	5.47	113	260165	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
50) Toluene-d8	8.70	98	899623	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
62) 4-Bromofluorobenzene	11.12	95	411880	56.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.50%	
Target Compounds						
20) Methylene Chloride	2.84	84	16266	2.270	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX073120\
Data File : VX017733.D
Acq On : 01 Aug 2020 04:25
Operator : JC/SP
Sample : PB130661ZHE#29
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130661ZHE#29

Quant Time: Aug 01 05:29:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

