

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073123\
 Data File : VX036725.D
 Acq On : 31 Jul 2023 17:11
 Operator : JC/MD
 Sample : 03808-03DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1710DL

Quant Time: Aug 01 04:39:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	127633	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	230702	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208976	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93851	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87133	46.138	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.280%
35) Dibromofluoromethane	5.391	113	73707	46.482	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.960%
50) Toluene-d8	8.647	98	273823	48.694	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.380%
62) 4-Bromofluorobenzene	11.079	95	101028	47.619	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.240%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4116	5.252	ug/l	96
52) Toluene	8.720	92	220396	51.151	ug/l	98
67) Ethyl Benzene	10.195	91	3753	0.446	ug/l	95
68) m/p-Xylenes	10.305	106	3478	1.108	ug/l	91
69) o-Xylene	10.640	106	3064	0.978	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073123\
Data File : VX036725.D
Acq On : 31 Jul 2023 17:11
Operator : JC/MD
Sample : 03808-03DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1710DL

Quant Time: Aug 01 04:39:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

