

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073123\
 Data File : VX036726.D
 Acq On : 31 Jul 2023 17:35
 Operator : JC/MD
 Sample : 03808-04DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1711DL

Quant Time: Aug 01 04:40:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	230962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213376	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94710	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85701	45.280	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.560%
35) Dibromofluoromethane	5.391	113	73452	46.269	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.540%
50) Toluene-d8	8.647	98	277274	49.252	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.500%
62) 4-Bromofluorobenzene	11.079	95	103645	48.798	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.600%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1501	1.911	ug/l #	80
52) Toluene	8.720	92	241374	55.957	ug/l	97
67) Ethyl Benzene	10.195	91	1806	0.210	ug/l	96
68) m/p-Xylenes	10.305	106	2314	0.722	ug/l	98
69) o-Xylene	10.640	106	827	0.258	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073123\
Data File : VX036726.D
Acq On : 31 Jul 2023 17:35
Operator : JC/MD
Sample : 03808-04DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1711DL

Quant Time: Aug 01 04:40:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

