

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073123\
 Data File : VX036727.D
 Acq On : 31 Jul 2023 17:59
 Operator : JC/MD
 Sample : 03808-05DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1712DL

Quant Time: Aug 01 04:40:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	229196	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92826	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84703	44.629	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.260%
35) Dibromofluoromethane	5.385	113	73187	46.457	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.920%
50) Toluene-d8	8.647	98	274063	49.057	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.120%
62) 4-Bromofluorobenzene	11.079	95	100597	47.728	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2130	2.705	ug/l	97
52) Toluene	8.720	92	410316	95.855	ug/l	97
68) m/p-Xylenes	10.299	106	742	0.235	ug/l	96
69) o-Xylene	10.640	106	737	0.234	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073123\
Data File : VX036727.D
Acq On : 31 Jul 2023 17:59
Operator : JC/MD
Sample : 03808-05DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1712DL

Quant Time: Aug 01 04:40:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

