

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073123\
 Data File : VX036729.D
 Acq On : 31 Jul 2023 18:47
 Operator : JC/MD
 Sample : 03808-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1714DL

Quant Time: Aug 01 04:41:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	125013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	223996	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202409	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84289	45.568	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.140%
35) Dibromofluoromethane	5.391	113	72379	47.011	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.020%
50) Toluene-d8	8.647	98	266492	48.809	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.079	95	96999	47.089	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.180%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2328	3.033	ug/l	97
25) 2-Butanone	4.574	43	3209	2.577	ug/l	99
51) 4-Methyl-2-Pentanone	8.580	43	724	0.293	ug/l	90
52) Toluene	8.720	92	642829	153.659	ug/l	98
68) m/p-Xylenes	10.299	106	1454	0.478	ug/l	97
69) o-Xylene	10.646	106	743	0.245	ug/l	91
84) 1,2,4-Trimethylbenzene	11.750	105	4544	0.745	ug/l	92
95) Naphthalene	13.774	128	3227	0.506	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073123\
Data File : VX036729.D
Acq On : 31 Jul 2023 18:47
Operator : JC/MD
Sample : 03808-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1714DL

Quant Time: Aug 01 04:41:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

