



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073123\
Data File : VX036707.D
Acq On : 31 Jul 2023 09:46
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Title : SW846 8260
Last Update : Fri Jul 21 12:14:12 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	10532	PASS
75	95	30	60	52.2	28947	PASS
95	95	100	100	100.0	55472	PASS
96	95	5	9	6.6	3668	PASS
173	174	0.00	2	0.8	338	PASS
174	95	50	100	73.8	40917	PASS
175	174	5	9	7.5	3074	PASS
176	174	95	101	95.4	39027	PASS
177	176	5	9	6.6	2589	PASS