

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073125\
 Data File : VX047195.D
 Acq On : 31 Jul 2025 11:56
 Operator : JC/MD
 Sample : PB169061TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB169061TB

Quant Time: Aug 01 01:48:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	20664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	34995	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	34220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	17086	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	1430	5.555	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	11.120%#
35) Dibromofluoromethane	5.410	113	710	3.286	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	6.580%#
50) Toluene-d8	8.653	98	2721	3.889	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	7.780%#
62) 4-Bromofluorobenzene	11.085	95	1860	6.168	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	12.340%#
Target Compounds						
					Qvalue	
3) Chloromethane	1.319	50	300	1.105	ug/l #	43
4) Vinyl Chloride	1.386	62	101	0.346	ug/l #	43
5) Bromomethane	1.642	94	981	6.188	ug/l	97
6) Chloroethane	1.691	64	78	0.407	ug/l #	43
8) Diethyl Ether	2.148	74	784	5.109	ug/l #	31
9) 1,1,2-Trichlorotrifluo...	2.343	101	112	0.427	ug/l #	20
10) Methyl Iodide	2.483	142	3220	12.050	ug/l	95
11) Tert butyl alcohol	2.953	59	362	1.819	ug/l #	72
12) 1,1-Dichloroethene	2.337	96	72	0.277	ug/l #	47
13) Acrolein	2.252	56	236	4.258	ug/l #	37
14) Allyl chloride	2.672	41	1252	2.667	ug/l #	48
15) Acrylonitrile	3.093	53	293	2.387	ug/l #	49
16) Acetone	2.386	43	53536	519.171	ug/l	97
17) Carbon Disulfide	2.532	76	3014	4.012	ug/l #	62
18) Methyl Acetate	2.715	43	29679	108.304	ug/l	96
20) Methylene Chloride	2.806	84	16083	56.692	ug/l	96
21) trans-1,2-Dichloroethene	3.111	96	161	0.606	ug/l #	68
25) 2-Butanone	4.599	43	5869	39.088	ug/l	94
28) Bromochloromethane	4.904	49	128	0.554	ug/l #	15
29) Tetrahydrofuran	5.019	42	150	1.548	ug/l #	35
36) 1,1-Dichloropropene	5.568	75	1213	3.293	ug/l #	44
37) Ethyl Acetate	4.739	43	15252	44.940	ug/l	99
38) Carbon Tetrachloride	5.556	117	1292	3.287	ug/l #	12
39) Methylcyclohexane	7.373	83	306	0.665	ug/l #	56
40) Benzene	6.056	78	629	0.590	ug/l #	1
41) Methacrylonitrile	5.013	41	65	0.366	ug/l #	1
43) Isopropyl Acetate	6.348	43	12921	24.421	ug/l #	91
44) Trichloroethene	7.147	130	95	0.348	ug/l	1
46) Dibromomethane	7.598	93	161	0.860	ug/l #	44
48) Methyl methacrylate	7.799	41	115213	399.727	ug/l #	46
49) 1,4-Dioxane	7.653	88	20	6.101	ug/l #	1
51) 4-Methyl-2-Pentanone	8.549	43	157954	502.267	ug/l #	49
52) Toluene	8.726	92	257	0.392	ug/l #	49
53) t-1,3-Dichloropropene	8.988	75	219	0.582	ug/l #	44
54) cis-1,3-Dichloropropene	8.549	75	70867	172.076	ug/l #	60
56) Ethyl methacrylate	9.007	69	88	0.242	ug/l #	22

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073125\
 Data File : VX047195.D
 Acq On : 31 Jul 2025 11:56
 Operator : JC/MD
 Sample : PB169061TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB169061TB


Quant Time: Aug 01 01:48:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

57) 1,3-Dichloropropane	9.482	76	4971	11.637	ug/l #	43
58) 2-Chloroethyl Vinyl ether	8.269	63	127	0.627	ug/l #	47
59) 2-Hexanone	9.482	43	70805	340.781	ug/l #	53
60) Dibromochloromethane	9.275	129	153	0.531	ug/l #	25
64) Tetrachloroethene	9.281	164	95	0.385	ug/l #	65
65) Chlorobenzene	10.079	112	313	0.390	ug/l #	4
67) Ethyl Benzene	10.195	91	485	0.346	ug/l	91
68) m/p-Xylenes	10.305	106	472	0.900	ug/l	81
69) o-Xylene	10.646	106	222	0.445	ug/l	50
70) Styrene	10.677	104	274	0.320	ug/l #	70
73) Isopropylbenzene	10.963	105	532	0.395	ug/l	100
74) N-amyl acetate	10.640	43	25204	46.341	ug/l #	53
76) 1,2,3-Trichloropropane	11.092	75	778	2.460	ug/l #	33
77) Bromobenzene	11.207	156	82	0.254	ug/l	75
78) n-propylbenzene	11.305	91	689	0.428	ug/l	85
79) 2-Chlorotoluene	11.372	91	705	0.732	ug/l	88
80) 1,3,5-Trimethylbenzene	11.451	105	993	0.892	ug/l	92
81) trans-1,4-Dichloro-2-b...	11.092	75	778	6.338	ug/l #	8
82) 4-Chlorotoluene	11.463	91	984	0.876	ug/l	98
83) tert-Butylbenzene	11.713	119	671	0.600	ug/l #	52
84) 1,2,4-Trimethylbenzene	11.750	105	1510	1.359	ug/l	77
85) sec-Butylbenzene	11.890	105	1782	1.276	ug/l	89
86) p-Isopropyltoluene	12.012	119	1186	1.024	ug/l	89
87) 1,3-Dichlorobenzene	11.969	146	676	1.100	ug/l	94
88) 1,4-Dichlorobenzene	12.036	146	788	1.266	ug/l	76
89) n-Butylbenzene	12.335	91	2186	2.013	ug/l	92
90) Hexachloroethane	12.536	117	177	0.854	ug/l	65
91) 1,2-Dichlorobenzene	12.341	146	506	0.862	ug/l	82
92) 1,2-Dibromo-3-Chloropr...	12.939	75	28	0.379	ug/l #	59
93) 1,2,4-Trichlorobenzene	13.591	180	1308	3.372	ug/l	85
94) Hexachlorobutadiene	13.719	225	1645	12.469	ug/l	94
95) Naphthalene	13.780	128	3227	2.838	ug/l #	92
96) 1,2,3-Trichlorobenzene	13.963	180	1128	3.018	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 01 01:48:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.568 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:168 Resp: 20664
Ion Ratio Lower Upper
168 100
99 65.3 48.0 72.0



#3
Chloromethane
Concen: 1.105 ug/l
RT: 1.319 min Scan# 38
Delta R.T. 0.013 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 50 Resp: 300
Ion Ratio Lower Upper
50 100
52 0.0 25.4 38.2#









#4
Vinyl Chloride
Concen: 0.346 ug/l
RT: 1.386 min Scan# 49
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 62 Resp: 101
Ion Ratio Lower Upper
62 100
64 0.0 25.2 37.8#



#5
Bromomethane
Concen: 6.188 ug/l
RT: 1.642 min Scan# 91
Delta R.T. 0.025 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 94 Resp: 981
Ion Ratio Lower Upper
94 100
96 87.6 72.3 108.5









#6
Chloroethane
Concen: 0.407 ug/l
RT: 1.691 min Scan# 99
Delta R.T. -0.012 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 64 Resp: 78
Ion Ratio Lower Upper
64 100
66 0.0 25.0 37.6#



#8
Diethyl Ether
Concen: 5.109 ug/l
RT: 2.148 min Scan# 174
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 74 Resp: 784
Ion Ratio Lower Upper
74 100
45 177.8 53.0 159.0#









#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.427 ug/l
RT: 2.343 min Scan# 207
Delta R.T. -0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:101 Resp: 112
Ion Ratio Lower Upper
101 100
85 0.0 35.4 53.0#
151 0.0 59.2 88.8#



#10
Methyl Iodide
Concen: 12.050 ug/l
RT: 2.483 min Scan# 229
Delta R.T. 0.013 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion:142 Resp: 3220
Ion Ratio Lower Upper
142 100
127 49.9 36.6 54.8
141 14.4 10.9 16.3





#11

Tert butyl alcohol

Concen: 1.819 ug/l

RT: 2.953 min Scan# 307

Delta R.T. -0.005 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56

Instrument :

MSVOA_X

ClientSampleId :





PB169061TB

Tgt Ion: 59 Resp: 362

Ion Ratio Lower Upper

59 100

57 0.0 8.5 12.7#



#12

1,1-Dichloroethene

Concen: 0.277 ug/l

RT: 2.337 min Scan# 205

Delta R.T. 0.000 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56




Tgt Ion: 96 Resp: 72


Ion Ratio Lower Upper

96 100

61 222.9 132.6 198.8#

98 0.0 50.2 75.4#









#13
Acrolein
Concen: 4.258 ug/l
RT: 2.252 min Scan# 191
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 56 Resp: 236
Ion Ratio Lower Upper
56 100
55 18.2 55.7 83.5#



#14
Allyl chloride
Concen: 2.667 ug/l
RT: 2.672 min Scan# 260
Delta R.T. -0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 41 Resp: 1252
Ion Ratio Lower Upper
41 100
39 18.1 52.6 78.8#
76 12.1 25.7 38.5#









#15
Acrylonitrile
Concen: 2.387 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.019 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 53 Resp: 293
Ion Ratio Lower Upper
53 100
52 39.9 65.7 98.5#
51 0.0 28.8 43.2#



#16
Acetone
Concen: 519.171 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 43 Resp: 53536
Ion Ratio Lower Upper
43 100
58 28.6 24.0 36.0









#17
Carbon Disulfide
Concen: 4.012 ug/l
RT: 2.532 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 76 Resp: 3014
Ion Ratio Lower Upper
76 100
78 22.4 7.0 10.4#



#18
Methyl Acetate
Concen: 108.304 ug/l
RT: 2.715 min Scan# 267
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 43 Resp: 29679
Ion Ratio Lower Upper
43 100
74 20.3 17.7 26.5









#20
Methylene Chloride
Concen: 56.692 ug/l
RT: 2.806 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:	84	Resp:	16083
Ion	Ratio	Lower	Upper
84	100		
49	134.7	101.6	152.4
51	37.9	30.3	45.5
86	64.2	50.6	75.8



#21
trans-1,2-Dichloroethene
Concen: 0.606 ug/l
RT: 3.111 min Scan# 332
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion:	96	Resp:	161
Ion	Ratio	Lower	Upper
96	100		
61	151.1	111.1	166.7
98	0.0	49.6	74.4#









#25
2-Butanone
Concen: 39.088 ug/l
RT: 4.599 min Scan# 51
Delta R.T. 0.037 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 43 Resp: 5869
Ion Ratio Lower Upper
43 100
72 26.4 18.8 28.2



#28
Bromochloromethane
Concen: 0.554 ug/l
RT: 4.904 min Scan# 626
Delta R.T. -0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 49 Resp: 128
Ion Ratio Lower Upper
49 100
129 0.0 0.0 3.8
130 0.0 59.0 88.4#









#29
Tetrahydrofuran
Concen: 1.548 ug/l
RT: 5.019 min Scan# 644
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 42 Resp: 150
Ion Ratio Lower Upper
42 100
72 0.0 33.7 50.5#
71 0.0 31.0 46.4#



#33
1,2-Dichloroethane-d4
Concen: 5.555 ug/l
RT: 5.970 min Scan# 801
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 65 Resp: 1430
Ion Ratio Lower Upper
65 100
67 42.0 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.775 min Scan# 91

Delta R.T. 0.006 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56

Instrument :

MSVOA_X

ClientSampleId :

PB169061TB





Tgt Ion:114 Resp: 34995

Ion Ratio Lower Upper

114 100

63 18.6 0.0 43.2

88 16.3 0.0 33.8



#35

Dibromofluoromethane

Concen: 3.286 ug/l

RT: 5.410 min Scan# 709

Delta R.T. 0.013 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56




Tgt Ion:113 Resp: 710


Ion Ratio Lower Upper

113 100

111 80.4 82.6 124.0#

192 0.0 14.9 22.3#









#36
1,1-Dichloropropene
Concen: 3.293 ug/l
RT: 5.568 min Scan# 71
Delta R.T. -0.134 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 75 Resp: 1213
Ion Ratio Lower Upper
75 100
110 2.1 17.2 51.5#
77 0.0 25.1 37.7#



#37
Ethyl Acetate
Concen: 44.940 ug/l
RT: 4.739 min Scan# 599
Delta R.T. 0.019 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56


Tgt Ion: 43 Resp: 15252
Ion Ratio Lower Upper
43 100
61 13.1 10.2 15.2
70 10.5 8.1 12.1








#38
Carbon Tetrachloride
Concen: 3.287 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.134 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Instrument :
MSVOA_X
ClientSampleId :
PB169061TB







Tgt Ion:117 Resp: 1292
Ion Ratio Lower Upper
117 100
119 0.0 76.6 115.0#
121 0.0 24.4 36.6#



#39
Methylcyclohexane
Concen: 0.665 ug/l
RT: 7.373 min Scan# 1031
Delta R.T. -0.012 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 83 Resp: 306
Ion Ratio Lower Upper
83 100
55 55.2 64.5 96.7#
98 0.0 36.9 55.3#









#40
Benzene
Concen: 0.590 ug/l
RT: 6.056 min Scan# 813
Delta R.T. 0.013 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 78 Resp: 629
Ion Ratio Lower Upper
78 100
77 129.9 19.0 28.4#



#41
Methacrylonitrile
Concen: 0.366 ug/l
RT: 5.013 min Scan# 644
Delta R.T. 0.086 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 41 Resp: 65
Ion Ratio Lower Upper
41 100
39 158.5 44.2 66.2#
67 0.0 56.4 84.6#
52 0.0 23.5 35.3#









#43
Isopropyl Acetate
Concen: 24.421 ug/l
RT: 6.348 min Scan# 863
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 43 Resp: 12921
Ion Ratio Lower Upper
43 100
61 12.2 15.1 22.7#
87 11.7 9.7 14.5



#44
Trichloroethene
Concen: 0.348 ug/l
RT: 7.147 min Scan# 994
Delta R.T. 0.019 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion:130 Resp: 95
Ion Ratio Lower Upper
130 100
95 0.0 0.0 198.6





#46

Dibromomethane

Concen: 0.860 ug/l

RT: 7.598 min Scan# 1100

Delta R.T. 0.013 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56

Instrument :

MSVOA_X

ClientSampleId :

PB169061TB





Tgt Ion: 93 Resp: 161

Ion Ratio Lower Upper

93 100

95 14.3 65.9 98.9#

174 57.1 74.6 112.0#



#48

Methyl methacrylate

Concen: 399.727 ug/l

RT: 7.799 min Scan# 1101

Delta R.T. 0.104 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56




Tgt Ion: 41 Resp: 115213


Ion Ratio Lower Upper

41 100

69 0.0 62.3 93.5#

39 53.3 44.8 67.2





#49

1,4-Dioxane

Concen: 6.101 ug/l

RT: 7.653 min Scan# 1078

Delta R.T. -0.006 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56

Instrument :





MSVOA_X

ClientSampleId :

PB169061TB

Tgt Ion: 88 Resp: 20

Ion	Ratio	Lower	Upper
88	100		
43	0.0	33.6	50.4#
58	1890.0	58.8	88.2#



#50

Toluene-d8

Concen: 3.889 ug/l

RT: 8.653 min Scan# 1241




Delta R.T. 0.000 min


Lab File: VX047195.D

Acq: 31 Jul 2025 11:56

Tgt Ion: 98 Resp: 2721

Ion	Ratio	Lower	Upper
98	100		
100	82.6	53.3	79.9#









#51
4-Methyl-2-Pentanone
Concen: 502.267 ug/l
RT: 8.549 min Scan# 11
Delta R.T. -0.024 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 43 Resp: 157954
Ion Ratio Lower Upper
43 100
58 7.4 30.6 45.8#



#52
Toluene
Concen: 0.392 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 92 Resp: 257
Ion Ratio Lower Upper
92 100
91 243.2 137.9 206.9#





#53

t-1,3-Dichloropropene

Concen: 0.582 ug/l

RT: 8.988 min Scan# 11

Delta R.T. 0.013 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56

Instrument :

MSVOA_X

ClientSampleId :





PB169061TB

Tgt Ion: 75 Resp: 219

Ion Ratio Lower Upper

75 100

77 0.0 24.5 36.7#



#54

cis-1,3-Dichloropropene

Concen: 172.076 ug/l

RT: 8.549 min Scan# 1224

Delta R.T. 0.183 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56




Tgt Ion: 75 Resp: 70867


Ion Ratio Lower Upper

75 100

77 0.5 25.4 38.0#

39 31.0 41.8 62.6#









#56
Ethyl methacrylate
Concen: 0.242 ug/l
RT: 9.007 min Scan# 11
Delta R.T. -0.109 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 69 Resp: 88
Ion Ratio Lower Upper
69 100
41 0.0 56.1 84.1#
39 0.0 31.1 46.7#



#57
1,3-Dichloropropane
Concen: 11.637 ug/l
RT: 9.482 min Scan# 1377
Delta R.T. 0.177 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 76 Resp: 4971
Ion Ratio Lower Upper
76 100
78 0.0 25.5 38.3#









#58
2-Chloroethyl Vinyl ether
Concen: 0.627 ug/l
RT: 8.269 min Scan# 1178
Delta R.T. 0.025 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 63 Resp: 127
Ion Ratio Lower Upper
63 100
106 0.0 22.2 33.4#



#59
2-Hexanone
Concen: 340.781 ug/l
RT: 9.482 min Scan# 1377
Delta R.T. 0.055 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 43 Resp: 70805
Ion Ratio Lower Upper
43 100
58 19.7 26.5 79.5#









#60
Dibromochloromethane
Concen: 0.531 ug/l
RT: 9.275 min Scan# 11
Delta R.T. -0.243 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:129 Resp: 153
Ion Ratio Lower Upper
129 100
127 12.4 38.6 115.8#



#62
4-Bromofluorobenzene
Concen: 6.168 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 95 Resp: 1860
Ion Ratio Lower Upper
95 100
174 68.5 0.0 144.6
176 62.9 0.0 139.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:117 Resp: 34220
Ion Ratio Lower Upper
117 100
82 58.0 45.4 68.2
119 31.0 26.1 39.1



#64
Tetrachloroethene
Concen: 0.385 ug/l
RT: 9.281 min Scan# 1344
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion:164 Resp: 95
Ion Ratio Lower Upper
164 100
166 126.6 106.2 159.4
129 79.8 76.6 114.8
131 0.0 74.4 111.6#









#65
Chlorobenzene
Concen: 0.390 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:112 Resp: 313
Ion Ratio Lower Upper
112 100
114 85.8 25.8 38.6#



#67
Ethyl Benzene
Concen: 0.346 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 91 Resp: 485
Ion Ratio Lower Upper
91 100
106 25.6 24.2 36.4









#68
m/p-Xylenes
Concen: 0.900 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:106 Resp: 472
Ion Ratio Lower Upper
106 100
91 239.0 167.4 251.0



#69
o-Xylene
Concen: 0.445 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion:106 Resp: 222
Ion Ratio Lower Upper
106 100
91 301.4 110.3 331.0





#70

Styrene

Concen: 0.320 ug/l

RT: 10.677 min Scan# 11

Delta R.T. 0.025 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56

Instrument :

MSVOA_X

ClientSampleId :

PB169061TB





Tgt Ion:104 Resp: 274

Ion Ratio Lower Upper

104 100

78 51.8 42.0 63.0

103 13.5 43.8 65.6#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.007 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56




Tgt Ion:152 Resp: 17086


Ion Ratio Lower Upper

152 100

115 60.6 45.6 136.7

150 160.7 0.0 354.6









#73
Isopropylbenzene
Concen: 0.395 ug/l
RT: 10.963 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:105 Resp: 532
Ion Ratio Lower Upper
105 100
120 25.9 12.9 38.7



#74
N-ethyl acetate
Concen: 46.341 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.201 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 43 Resp: 25204
Ion Ratio Lower Upper
43 100
70 1.4 32.3 48.5#
55 4.5 18.9 28.3#
61 9.9 18.7 28.1#









#76
1,2,3-Trichloropropane
Concen: 2.460 ug/l
RT: 11.092 min Scan# 1641
Delta R.T. -0.146 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 75 Resp: 778
Ion Ratio Lower Upper
75 100
77 0.0 21.6 65.0#



#77
Bromobenzene
Concen: 0.254 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.013 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion:156 Resp: 82
Ion Ratio Lower Upper
156 100
77 157.3 87.3 261.9
158 51.2 48.9 146.6









#78
n-propylbenzene
Concen: 0.428 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 91 Resp: 689
Ion Ratio Lower Upper
91 100
120 29.2 11.0 33.0



#79
2-Chlorotoluene
Concen: 0.732 ug/l
RT: 11.372 min Scan# 1687
Delta R.T. 0.013 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 91 Resp: 705
Ion Ratio Lower Upper
91 100
126 26.2 16.4 49.4









#80
1,3,5-Trimethylbenzene
Concen: 0.892 ug/l
RT: 11.451 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:105 Resp: 993
Ion Ratio Lower Upper
105 100
120 53.3 23.8 71.5



#81
trans-1,4-Dichloro-2-butene
Concen: 6.338 ug/l
RT: 11.092 min Scan# 1641
Delta R.T. 0.074 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 75 Resp: 778
Ion Ratio Lower Upper
75 100
53 0.0 84.2 126.4#
89 0.0 36.0 54.0#









#82
4-Chlorotoluene
Concen: 0.876 ug/l
RT: 11.463 min Scan# 11
Delta R.T. 0.013 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion: 91 Resp: 984
Ion Ratio Lower Upper
91 100
126 27.1 14.1 42.3



#83
tert-Butylbenzene
Concen: 0.600 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion: 119 Resp: 671
Ion Ratio Lower Upper
119 100
91 100.7 29.6 88.8#
134 8.0 11.4 34.1#









#84
1,2,4-Trimethylbenzene
Concen: 1.359 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:105 Resp: 1510
Ion Ratio Lower Upper
105 100
120 28.9 22.0 66.0



#85
sec-Butylbenzene
Concen: 1.276 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion:105 Resp: 1782
Ion Ratio Lower Upper
105 100
134 14.7 9.9 29.7





#86

p-Isopropyltoluene

Concen: 1.024 ug/l

RT: 12.012 min Scan# 1186

Delta R.T. 0.006 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56

Instrument :

MSVOA_X

ClientSampleId :

PB169061TB





Tgt Ion:119 Resp: 1186

Ion Ratio Lower Upper

119 100

134 28.8 12.8 38.4

91 33.3 12.4 37.4



#87

1,3-Dichlorobenzene

Concen: 1.100 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56




Tgt Ion:146 Resp: 676


Ion Ratio Lower Upper

146 100

111 39.9 21.0 63.0

148 58.3 32.0 96.0





#88

1,4-Dichlorobenzene

Concen: 1.266 ug/l

RT: 12.036 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56

Instrument :

MSVOA_X

ClientSampleId :

PB169061TB





Tgt Ion:146 Resp: 788

Ion Ratio Lower Upper

146 100

111 61.5 20.6 61.9

148 51.1 33.0 98.9



#89

n-Butylbenzene

Concen: 2.013 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.007 min

Lab File: VX047195.D

Acq: 31 Jul 2025 11:56




Tgt Ion: 91 Resp: 2186


Ion Ratio Lower Upper

91 100

92 48.2 27.5 82.5

134 28.3 12.8 38.3









#90
Hexachloroethane
Concen: 0.854 ug/l
RT: 12.536 min Scan# 117
Delta R.T. 0.000 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:117 Resp: 177
Ion Ratio Lower Upper
117 100
201 40.7 34.8 104.5



#91
1,2-Dichlorobenzene
Concen: 0.862 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion:146 Resp: 506
Ion Ratio Lower Upper
146 100
111 46.0 21.6 64.6
148 84.6 31.9 95.7









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 0.379 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. 0.000 min
 Lab File: VX047195.D
 Acq: 31 Jul 2025 11:56




Instrument :
 MSVOA_X
 ClientSampleId :
 PB169061TB


Tgt Ion: 75 Resp: 28
 Ion Ratio Lower Upper
 75 100
 155 0.0 41.3 124.1#
 157 107.1 52.8 158.3



#93
 1,2,4-Trichlorobenzene
 Concen: 3.372 ug/l
 RT: 13.591 min Scan# 2051
 Delta R.T. 0.006 min
 Lab File: VX047195.D
 Acq: 31 Jul 2025 11:56

Tgt Ion: 180 Resp: 1308
 Ion Ratio Lower Upper
 180 100
 182 75.5 47.3 141.8
 145 34.3 16.4 49.1









#94
Hexachlorobutadiene
Concen: 12.469 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56




Instrument :
MSVOA_X
ClientSampleId :
PB169061TB


Tgt Ion:	225	Resp:	1645
Ion Ratio	Lower	Upper	
225	100		
223	64.7	31.6	95.0
227	58.5	32.9	98.6



#95
Naphthalene
Concen: 2.838 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.007 min
Lab File: VX047195.D
Acq: 31 Jul 2025 11:56

Tgt Ion:	128	Resp:	3227
Ion Ratio	Lower	Upper	
128	100		
127	7.8	10.1	15.1#
129	9.2	8.6	13.0





#96

1,2,3-Trichlorobenzene

Concen: 3.018 ug/l

RT: 13.963 min Scan# 21

Delta R.T. 0.007 min

Lab File: VX047195.D


Acq: 31 Jul 2025 11:56

Instrument :

MSVOA_X

ClientSampleId :

PB169061TB



Tgt Ion:180 Resp: 1128

Ion Ratio Lower Upper

180 100

182 106.6 46.9 140.6

145 42.2 17.3 52.0

