

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073125\
 Data File : VX047191.D
 Acq On : 31 Jul 2025 10:19
 Operator : JC/MD
 Sample : VX0731MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0731MBL01

Quant Time: Aug 01 01:46:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	243708	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	454704	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	438363	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	233519	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	165900	54.647	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.300%
35) Dibromofluoromethane	5.397	113	136162	48.500	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.000%
50) Toluene-d8	8.647	98	503680	55.398	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	110.800%
62) 4-Bromofluorobenzene	11.079	95	212310	54.186	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073125\
Data File : VX047191.D
Acq On : 31 Jul 2025 10:19
Operator : JC/MD
Sample : VX0731MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0731MBL01

Quant Time: Aug 01 01:46:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

