

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073125\
 Data File : VX047197.D
 Acq On : 31 Jul 2025 12:39
 Operator : JC/MD
 Sample : PB169061TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB169061TB

Quant Time: Aug 01 01:50:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	351350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	653721	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	605698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	306388	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	248783	56.842	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 113.680%		
35) Dibromofluoromethane	5.397	113	208691	51.704	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.400%		
50) Toluene-d8	8.653	98	720453	55.117	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 110.240%		
62) 4-Bromofluorobenzene	11.079	95	308912	54.839	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 109.680%		
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	55818	31.836	ug/l	98
18) Methyl Acetate	2.715	43	31614	6.785	ug/l	100
20) Methylene Chloride	2.806	84	16708	3.464	ug/l	92
37) Ethyl Acetate	4.733	43	20647	3.257	ug/l #	96
43) Isopropyl Acetate	6.361	43	14641	1.481	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX073125\
Data File : VX047197.D
Acq On : 31 Jul 2025 12:39
Operator : JC/MD
Sample : PB169061TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB169061TB

Quant Time: Aug 01 01:50:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

