

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080118\
 Data File : VX003798.D
 Acq On : 01 Aug 2018 17:08
 Operator : JC/SP
 Sample : J4143-05
 Misc : 11.19G/5.0mL/MSVOA X/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-01(3-5)

Manual Integrations
 APPROVED

apatel
 8/2/2018 4:20:01 PM

Quant Time: Aug 02 06:48:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X073018S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 31 11:17:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	145904	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	215935	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	177976	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	69902	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	58306	39.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.02%	
35) Dibromofluoromethane	5.51	113	41938	32.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.96%	
50) Toluene-d8	8.72	98	197566	37.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.54%	
62) 4-Bromofluorobenzene	11.14	95	59320	30.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.70%	
Target Compounds						
16) Acetone	2.45	43	31800	56.418	ug/l	95
17) Carbon Disulfide	2.57	76	22267	5.248	ug/l	99
40) Benzene	6.15	78	6553	1.131	ug/l #	90
80) 1,3,5-Trimethylbenzene	11.51	105	6576m	1.761	ug/l	
84) 1,2,4-Trimethylbenzene	11.81	105	5563	1.447	ug/l	98
95) Naphthalene	13.83	128	18772	5.152	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080118\
Data File : VX003798.D
Acq On : 01 Aug 2018 17:08
Operator : JC/SP
Sample : J4143-05
Misc : 11.19G/5.0mL/MSVOA X/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-01(3-5)

Manual Integrations
APPROVED

apatel
8/2/2018 4:20:01 PM

Quant Time: Aug 02 06:48:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X073018S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 31 11:17:41 2018
Response via : Initial Calibration

