

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080119\
 Data File : VX011271.D
 Acq On : 01 Aug 2019 20:30
 Operator : JC/SP
 Sample : K4118-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-MW127S-20190730

Quant Time: Aug 02 02:27:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	143604	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	230394	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	215798	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	99547	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	74650	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	5.49	113	72186	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	8.71	98	279014	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
62) 4-Bromofluorobenzene	11.13	95	92296	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
Target Compounds						
3) Chloromethane	1.32	50	1686	1.354	ug/l #	83
16) Acetone	2.43	43	3139	4.560	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080119\
Data File : VX011271.D
Acq On : 01 Aug 2019 20:30
Operator : JC/SP
Sample : K4118-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-MW127S-20190730

Quant Time: Aug 02 02:27:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

