

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080119\
 Data File : VX011274.D
 Acq On : 01 Aug 2019 21:40
 Operator : JC/SP
 Sample : K4118-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-MW126D-20190730

Quant Time: Aug 02 02:35:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	139638	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	223615	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	209317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	96994	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	72536	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
35) Dibromofluoromethane	5.49	113	70528	48.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.62%	
50) Toluene-d8	8.71	98	270549	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
62) 4-Bromofluorobenzene	11.13	95	90557	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
Target Compounds						
16) Acetone	2.43	43	2084	3.114	ug/l #	Qvalue 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080119\
Data File : VX011274.D
Acq On : 01 Aug 2019 21:40
Operator : JC/SP
Sample : K4118-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-MW126D-20190730

Quant Time: Aug 02 02:35:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

