

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080119\
 Data File : VX011286.D
 Acq On : 02 Aug 2019 02:21
 Operator : JC/SP
 Sample : K4118-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SA-PZ157I-20190731

Quant Time: Aug 02 03:59:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	129253	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	208000	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	197473	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	90483	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	68896	52.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.86%	
35) Dibromofluoromethane	5.49	113	66759	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
50) Toluene-d8	8.71	98	254543	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
62) 4-Bromofluorobenzene	11.13	95	83553	45.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.30%	
Target Compounds						
16) Acetone	2.43	43	3010	4.858	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX080119\
Data File : VX011286.D
Acq On : 02 Aug 2019 02:21
Operator : JC/SP
Sample : K4118-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SA-PZ157I-20190731

Quant Time: Aug 02 03:59:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

