

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080119\
 Data File : VX011289.D
 Acq On : 02 Aug 2019 03:31
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 04:06:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.386	0.401	-3.9	103	0.00
3 P	Chloromethane	0.434	0.455	-4.8	106	0.00
4 C	Vinyl Chloride	0.477	0.507	-6.3#	104	0.00
5 T	Bromomethane	0.340	0.357	-5.0	105	0.00
6 T	Chloroethane	0.335	0.365	-9.0	113	0.00
7 T	Trichlorofluoromethane	0.848	0.907	-7.0	103	0.00
8 T	Diethyl Ether	0.303	0.321	-5.9	108	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.459	0.465	-1.3	100	0.00
10 T	Methyl Iodide	0.579	0.653	-12.8	101	0.00
11 T	Tert butyl alcohol	0.106	0.117	-10.4	111	0.00
12 CM	1,1-Dichloroethene	0.455	0.470	-3.3#	102	0.00
13 T	Acrolein	0.066	0.065	1.5	102	0.00
14 T	Allyl chloride	0.606	0.665	-9.7	102	0.00
15 T	Acrylonitrile	0.240	0.270	-12.5	107	0.00
16 T	Acetone	0.240	0.224	6.7	93	0.00
17 T	Carbon Disulfide	1.068	1.185	-11.0	106	0.00
18 T	Methyl Acetate	0.548	0.603	-10.0	109	0.00
19 T	Methyl tert-butyl Ether	1.341	1.492	-11.3	106	0.00
20 T	Methylene Chloride	0.525	0.542	-3.2	105	0.00
21 T	trans-1,2-Dichloroethene	0.481	0.508	-5.6	101	0.00
22 T	Diisopropyl ether	1.333	1.461	-9.6	108	0.00
23 T	Vinyl Acetate	1.106	1.274	-15.2	105	0.00
24 P	1,1-Dichloroethane	0.786	0.845	-7.5	105	0.00
25 T	2-Butanone	0.341	0.369	-8.2	103	0.00
26 T	2,2-Dichloropropane	0.611	0.454	25.7#	70	0.00
27 T	cis-1,2-Dichloroethene	0.554	0.593	-7.0	106	0.00
28 T	Bromochloromethane	0.358	0.369	-3.1	106	0.00
29 T	Tetrahydrofuran	0.209	0.237	-13.4	107	0.00
30 C	Chloroform	0.854	0.927	-8.5#	107	0.00
31 T	Cyclohexane	0.648	0.687	-6.0	102	0.00
32 T	1,1,1-Trichloroethane	0.732	0.793	-8.3	104	0.00
33 S	1,2-Dichloroethane-d4	0.504	0.515	-2.2	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	99	0.00
35 S	Dibromofluoromethane	0.326	0.325	0.3	102	0.00
36 T	1,1-Dichloropropene	0.437	0.445	-1.8	102	0.00
37 T	Ethyl Acetate	0.438	0.477	-8.9	109	0.00
38 T	Carbon Tetrachloride	0.439	0.472	-7.5	103	0.00
39 T	Methylcyclohexane	0.517	0.513	0.8	100	0.00
40 TM	Benzene	1.303	1.384	-6.2	104	0.00
41 T	Methacrylonitrile	0.239	0.265	-10.9	109	0.00
42 TM	1,2-Dichloroethane	0.441	0.474	-7.5	108	0.00
43 T	Isopropyl Acetate	0.700	0.757	-8.1	107	0.00
44 TM	Trichloroethene	0.394	0.407	-3.3	103	0.00
45 C	1,2-Dichloropropane	0.337	0.348	-3.3#	104	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080119\
 Data File : VX011289.D
 Acq On : 02 Aug 2019 03:31
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 04:06:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.246	0.259	-5.3	107	0.00
47 T	Bromodichloromethane	0.417	0.464	-11.3	108	0.00
48 T	Methyl methacrylate	0.329	0.375	-14.0	111	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	105	0.00
50 S	Toluene-d8	1.232	1.238	-0.5	104	0.00
51 T	4-Methyl-2-Pentanone	0.452	0.516	-14.2	109	0.00
52 CM	Toluene	0.861	0.919	-6.7#	104	0.00
53 T	t-1,3-Dichloropropene	0.423	0.467	-10.4	100	0.00
54 T	cis-1,3-Dichloropropene	0.487	0.526	-8.0	100	0.00
55 T	1,1,2-Trichloroethane	0.354	0.385	-8.8	107	0.00
56 T	Ethyl methacrylate	0.505	0.576	-14.1	108	0.00
57 T	1,3-Dichloropropane	0.560	0.597	-6.6	107	0.00
58 T	2-Chloroethyl Vinyl ether	0.248	0.256	-3.2	100	0.00
59 T	2-Hexanone	0.350	0.404	-15.4	107	0.00
60 T	Dibromochloromethane	0.363	0.411	-13.2	107	0.00
61 T	1,2-Dibromoethane	0.383	0.416	-8.6	106	0.00
62 S	4-Bromofluorobenzene	0.440	0.453	-3.0	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.425	0.436	-2.6	107	0.00
65 PM	Chlorobenzene	1.072	1.105	-3.1	106	0.00
66 T	1,1,1,2-Tetrachloroethane	0.367	0.407	-10.9	106	0.00
67 C	Ethyl Benzene	1.795	1.910	-6.4#	106	0.00
68 T	m/p-Xylenes	0.694	0.748	-7.8	105	0.00
69 T	o-Xylene	0.682	0.723	-6.0	105	0.00
70 T	Styrene	1.129	1.253	-11.0	106	0.00
71 P	Bromoform	0.285	0.333	-16.8	106	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.420	3.613	-5.6	105	0.00
74 T	N-amyl acetate	1.236	1.409	-14.0	105	0.00
75 P	1,1,2,2-Tetrachloroethane	1.140	1.204	-5.6	105	0.00
76 T	1,2,3-Trichloropropane	0.911	0.997	-9.4	106	0.00
77 T	Bromobenzene	0.948	1.005	-6.0	106	0.00
78 T	n-propylbenzene	3.803	4.051	-6.5	105	0.00
79 T	2-Chlorotoluene	2.267	2.373	-4.7	106	0.00
80 T	1,3,5-Trimethylbenzene	2.863	3.053	-6.6	106	0.00
81 T	trans-1,4-Dichloro-2-butene	0.312	0.308	1.3	93	0.00
82 T	4-Chlorotoluene	2.642	2.772	-4.9	105	0.00
83 T	tert-Butylbenzene	2.808	2.960	-5.4	103	0.00
84 T	1,2,4-Trimethylbenzene	2.850	3.072	-7.8	104	0.00
85 T	sec-Butylbenzene	3.373	3.539	-4.9	103	0.00
86 T	p-Isopropyltoluene	3.103	3.341	-7.7	105	0.00
87 T	1,3-Dichlorobenzene	1.675	1.746	-4.2	103	0.00
88 T	1,4-Dichlorobenzene	1.728	1.765	-2.1	103	0.00
89 T	n-Butylbenzene	2.674	2.827	-5.7	102	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX080119\
Data File : VX011289.D
Acq On : 02 Aug 2019 03:31
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 02 04:06:03 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.459	0.507	-10.5	102	0.00
91 T	1,2-Dichlorobenzene	1.683	1.758	-4.5	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.221	0.252	-14.0	108	0.00
93 T	1,2,4-Trichlorobenzene	1.116	1.184	-6.1	102	0.00
94 T	Hexachlorobutadiene	0.566	0.563	0.5	98	0.00
95 T	Naphthalene	3.265	3.687	-12.9	105	0.00
96 T	1,2,3-Trichlorobenzene	1.135	1.189	-4.8	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6