


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080119\
 Data File : VX011258.D
 Acq On : 01 Aug 2019 12:25
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Title : SW846 8260
 Last Update : Fri Aug 02 01:55:00 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.4	3927	PASS
75	95	30	60	46.8	11210	PASS
95	95	100	100	100.0	23930	PASS
96	95	5	9	6.3	1509	PASS
173	174	0.00	2	1.2	234	PASS
174	95	50	100	78.6	18815	PASS
175	174	5	9	7.7	1441	PASS
176	174	95	101	98.5	18531	PASS
177	176	5	9	6.1	1132	PASS