


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX080119\
Data File : VX011266.D
Acq On : 01 Aug 2019 17:46
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Title : SW846 8260
Last Update : Fri Aug 02 01:55:00 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	5081	PASS
75	95	30	60	50.5	14755	PASS
95	95	100	100	100.0	29234	PASS
96	95	5	9	7.6	2209	PASS
173	174	0.00	2	1.5	396	PASS
174	95	50	100	91.9	26869	PASS
175	174	5	9	8.2	2191	PASS
176	174	95	101	97.0	26061	PASS
177	176	5	9	6.6	1720	PASS