

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030318.D
 Acq On : 01 Aug 2022 13:29
 Operator : JC/MD
 Sample : N3773-18ME
 Misc : 6.50g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DL-3(14.5-15)ME

Quant Time: Aug 02 01:51:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	159644	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.763	114	274179	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	276348	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153107	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165846	51.686	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.380%
35) Dibromofluoromethane	5.391	113	128639	49.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.160%
50) Toluene-d8	8.653	98	509834	50.229	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.460%
62) 4-Bromofluorobenzene	11.085	95	174748	46.389	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.780%
Target Compounds						
						Qvalue
25) 2-Butanone	4.605	43	34123	12.361	ug/l	95
29) Tetrahydrofuran	5.056	42	2500	1.346	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030318.D
Acq On : 01 Aug 2022 13:29
Operator : JC/MD
Sample : N3773-18ME
Misc : 6.50g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DL-3(14.5-15)ME

Quant Time: Aug 02 01:51:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

