

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030321.D
 Acq On : 01 Aug 2022 14:44
 Operator : JC/MD
 Sample : N3861-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11R

Quant Time: Aug 02 01:52:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	153748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	269904	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	276978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153118	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165876	53.677	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.360%		
35) Dibromofluoromethane	5.398	113	132635	51.406	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.820%		
50) Toluene-d8	8.653	98	505887	50.629	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.260%		
62) 4-Bromofluorobenzene	11.086	95	175754	47.395	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 94.800%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	12230	16.225	ug/l #	83
16) Acetone	2.380	43	6639	4.196	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	7566	0.920	ug/l #	83
95) Naphthalene	13.780	128	18853	1.554	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030321.D
Acq On : 01 Aug 2022 14:44
Operator : JC/MD
Sample : N3861-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11R

Quant Time: Aug 02 01:52:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

