

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030324.D
 Acq On : 01 Aug 2022 15:54
 Operator : JC/MD
 Sample : N3861-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-25

Quant Time: Aug 02 01:54:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75231	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	146003	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	140928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127795	84.515	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 169.040%#	
35) Dibromofluoromethane	5.397	113	93411	66.927	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 133.860%#	
50) Toluene-d8	8.653	98	270221	49.994	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.980%	
62) 4-Bromofluorobenzene	11.085	95	90579	45.155	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.300%	
Target Compounds						Qvalue
16) Acetone	2.379	43	6245	8.067	ug/l	91
17) Carbon Disulfide	2.514	76	3753	1.183	ug/l	98
95) Naphthalene	13.780	128	2723	0.478	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030324.D
Acq On : 01 Aug 2022 15:54
Operator : JC/MD
Sample : N3861-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-25

Quant Time: Aug 02 01:54:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

