

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030325.D
 Acq On : 01 Aug 2022 16:18
 Operator : JC/MD
 Sample : N3861-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-072122

Quant Time: Aug 02 01:54:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	257662	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	263265	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	157723	54.960	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.920%	
35) Dibromofluoromethane	5.391	113	125747	51.052	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.100%	
50) Toluene-d8	8.653	98	478718	50.186	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.380%	
62) 4-Bromofluorobenzene	11.085	95	164616	46.501	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.000%	
Target Compounds						
16) Acetone	2.380	43	8087	5.504	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030325.D
Acq On : 01 Aug 2022 16:18
Operator : JC/MD
Sample : N3861-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-072122

Quant Time: Aug 02 01:54:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

