

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030330.D
 Acq On : 01 Aug 2022 18:15
 Operator : JC/MD
 Sample : N3964-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HMGP-WATER-1DL

Quant Time: Aug 02 01:55:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	144659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247376	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253165	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142727	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	143787	49.453	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.900%
35) Dibromofluoromethane	5.385	113	115237	48.730	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	8.653	98	434063	47.397	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.800%
62) 4-Bromofluorobenzene	11.085	95	150469	44.272	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.540%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2999	2.015	ug/l #	78
19) Methyl tert-butyl Ether	3.117	73	50489	6.522	ug/l	100
37) Ethyl Acetate	4.715	43	815208	169.154	ug/l	98
68) m/p-Xylenes	10.299	106	3082	0.662	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	2908	0.312	ug/l	95
95) Naphthalene	13.774	128	147041	13.003	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030330.D
Acq On : 01 Aug 2022 18:15
Operator : JC/MD
Sample : N3964-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HMGP-WATER-1DL

Quant Time: Aug 02 01:55:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

