

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030331.D
 Acq On : 01 Aug 2022 18:38
 Operator : JC/MD
 Sample : N3861-14 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2-INCH

Quant Time: Aug 02 01:56:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133099	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231186	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136217	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	151255	56.540	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 113.080%	
35) Dibromofluoromethane	5.391	113	121624	55.033	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.060%	
50) Toluene-d8	8.653	98	449659	52.539	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.080%	
62) 4-Bromofluorobenzene	11.085	95	156848	49.380	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.760%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.476	56	6793	1.749	ug/l #	88
39) Methylcyclohexane	7.385	83	3448	0.840	ug/l	91
40) Benzene	6.037	78	3083	0.308	ug/l	95
73) Isopropylbenzene	10.963	105	13762	1.294	ug/l	100
78) n-propylbenzene	11.305	91	40992	3.230	ug/l	99
95) Naphthalene	13.780	128	5732	0.531	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030331.D
Acq On : 01 Aug 2022 18:38
Operator : JC/MD
Sample : N3861-14 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2-INCH

Quant Time: Aug 02 01:56:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

