

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030333.D
 Acq On : 01 Aug 2022 19:25
 Operator : JC/MD
 Sample : N3964-05 10X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HMGP-SOIL-C

Quant Time: Aug 02 01:56:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	234977	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138193	51.442	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.880%
35) Dibromofluoromethane	5.385	113	107982	48.072	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.140%
50) Toluene-d8	8.653	98	416340	47.861	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.720%
62) 4-Bromofluorobenzene	11.085	95	152378	47.199	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.400%
Target Compounds						
					Qvalue	
68) m/p-Xylenes	10.305	106	6110	1.374	ug/l	88
95) Naphthalene	13.780	128	449986	39.201	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030333.D
 Acq On : 01 Aug 2022 19:25
 Operator : JC/MD
 Sample : N3964-05 10X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HMGP-SOIL-C

Quant Time: Aug 02 01:56:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

