

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030334.D
 Acq On : 01 Aug 2022 19:48
 Operator : JC/MD
 Sample : N3964-03 10X
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HMGP-SOIL-B

Quant Time: Aug 02 01:57:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135460	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238723	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	244400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	144372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138996	51.051	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 102.100%			
35) Dibromofluoromethane	5.391	113	109822	48.124	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.240%			
50) Toluene-d8	8.653	98	418061	47.305	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 94.600%			
62) 4-Bromofluorobenzene	11.085	95	154499	47.105	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.220%			
Target Compounds						
25) 2-Butanone	4.581	43	3904	1.667	ug/l	97
95) Naphthalene	13.780	128	267028	23.344	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030334.D
Acq On : 01 Aug 2022 19:48
Operator : JC/MD
Sample : N3964-03 10X
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HMGP-SOIL-B

Quant Time: Aug 02 01:57:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

