

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030335.D
 Acq On : 01 Aug 2022 20:12
 Operator : JC/MD
 Sample : N3964-01 10X
 Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HMGP-SOIL-A

Quant Time: Aug 02 01:57:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135413	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	234448	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	238107	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	150401	55.260	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.520%
35) Dibromofluoromethane	5.391	113	116906	52.162	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.320%
50) Toluene-d8	8.653	98	447942	51.610	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.220%
62) 4-Bromofluorobenzene	11.085	95	158663	49.257	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.520%
Target Compounds						
						Qvalue
95) Naphthalene	13.780	128	15928	1.463	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030335.D
Acq On : 01 Aug 2022 20:12
Operator : JC/MD
Sample : N3964-01 10X
Misc : 5.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HMGP-SOIL-A

Quant Time: Aug 02 01:57:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

