

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030336.D
 Acq On : 01 Aug 2022 20:35
 Operator : JC/MD
 Sample : N3861-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-25

Quant Time: Aug 02 01:57:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128251	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226057	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233465	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156212	60.600	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	121.200%
35) Dibromofluoromethane	5.391	113	122950	56.895	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.800%
50) Toluene-d8	8.653	98	457413	54.657	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.320%
62) 4-Bromofluorobenzene	11.085	95	159180	51.252	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.500%
Target Compounds						
16) Acetone	2.379	43	4321	3.274	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030336.D
Acq On : 01 Aug 2022 20:35
Operator : JC/MD
Sample : N3861-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-25

Quant Time: Aug 02 01:57:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

