

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030338.D
 Acq On : 01 Aug 2022 21:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 01:58:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.820	0.811	1.1	83	0.00
3 P	Chloromethane	0.811	0.677	16.5	75	0.00
4 C	Vinyl Chloride	0.819	0.747	8.8#	82	0.00
5 T	Bromomethane	0.355	0.204	42.5#	61	0.00
6 T	Chloroethane	0.470	0.508	-8.1	99	0.00
7 T	Trichlorofluoromethane	1.079	1.108	-2.7	92	0.00
8 T	Diethyl Ether	0.422	0.435	-3.1	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.799	0.809	-1.3	94	0.00
10 T	Methyl Iodide	0.696	0.302	56.6#	32#	0.00
11 T	Tert butyl alcohol	0.245	0.281	-14.7	103	0.02
12 CM	1,1-Dichloroethene	0.781	0.797	-2.0#	91	0.00
13 T	Acrolein	0.060	0.055	8.3	90	0.00
14 T	Allyl chloride	1.612	1.641	-1.8	94	0.00
15 T	Acrylonitrile	0.606	0.646	-6.6	94	0.00
16 T	Acetone	0.515	0.551	-7.0	97	0.00
17 T	Carbon Disulfide	2.108	1.817	13.8	79	0.00
18 T	Methyl Acetate	1.522	1.699	-11.6	105	0.00
19 T	Methyl tert-butyl Ether	2.676	2.920	-9.1	95	0.00
20 T	Methylene Chloride	1.184	0.958	19.1	90	0.00
21 T	trans-1,2-Dichloroethene	0.862	0.852	1.2	89	0.00
22 T	Diisopropyl ether	3.050	3.432	-12.5	97	0.00
23 T	Vinyl Acetate	2.645	2.951	-11.6	93	0.00
24 P	1,1-Dichloroethane	1.620	1.725	-6.5	96	0.00
25 T	2-Butanone	0.865	0.943	-9.0	95	0.00
26 T	2,2-Dichloropropane	1.172	1.068	8.9	86	0.00
27 T	cis-1,2-Dichloroethene	1.014	1.058	-4.3	94	0.00
28 T	Bromochloromethane	0.685	0.758	-10.7	103	0.00
29 T	Tetrahydrofuran	0.582	0.621	-6.7	94	0.00
30 C	Chloroform	1.604	1.720	-7.2#	96	0.00
31 T	Cyclohexane	1.459	1.464	-0.3	89	0.00
32 T	1,1,1-Trichloroethane	1.304	1.385	-6.2	94	0.00
33 S	1,2-Dichloroethane-d4	1.005	1.030	-2.5	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.478	0.488	-2.1	91	0.00
36 T	1,1-Dichloropropene	0.688	0.683	0.7	90	0.00
37 T	Ethyl Acetate	0.974	1.007	-3.4	95	0.00
38 T	Carbon Tetrachloride	0.642	0.647	-0.8	91	0.00
39 T	Methylcyclohexane	0.888	0.831	6.4	88	0.00
40 TM	Benzene	2.164	2.200	-1.7	92	0.00
41 T	Methacrylonitrile	0.522	0.561	-7.5	96	0.00
42 TM	1,2-Dichloroethane	0.748	0.777	-3.9	94	0.00
43 T	Isopropyl Acetate	1.428	1.550	-8.5	97	0.00
44 TM	Trichloroethene	0.591	0.558	5.6	90	0.00
45 C	1,2-Dichloropropane	0.577	0.604	-4.7#	96	0.00
46 T	Dibromomethane	0.393	0.392	0.3	93	0.00
47 T	Bromodichloromethane	0.714	0.757	-6.0	96	0.00
48 T	Methyl methacrylate	0.703	0.757	-7.7	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
 Data File : VX030338.D
 Acq On : 01 Aug 2022 21:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 02 01:58:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.016	0.016	0.0	94	0.00
50 S	Toluene-d8	1.851	1.808	2.3	85	0.00
51 T	4-Methyl-2-Pentanone	0.969	1.047	-8.0	96	0.00
52 CM	Toluene	1.320	1.325	-0.4#	89	0.00
53 T	t-1,3-Dichloropropene	0.769	0.784	-2.0	90	0.00
54 T	cis-1,3-Dichloropropene	0.838	0.871	-3.9	92	0.00
55 T	1,1,2-Trichloroethane	0.562	0.570	-1.4	92	0.00
56 T	Ethyl methacrylate	0.889	0.958	-7.8	93	0.00
57 T	1,3-Dichloropropane	0.928	0.960	-3.4	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.441	0.497	-12.7	93	0.00
59 T	2-Hexanone	0.765	0.825	-7.8	94	0.00
60 T	Dibromochloromethane	0.520	0.543	-4.4	91	0.00
61 T	1,2-Dibromoethane	0.587	0.591	-0.7	91	0.00
62 S	4-Bromofluorobenzene	0.687	0.694	-1.0	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.538	0.506	5.9	84	0.00
65 PM	Chlorobenzene	1.358	1.373	-1.1	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.462	0.501	-8.4	92	0.00
67 C	Ethyl Benzene	2.659	2.558	3.8#	91	0.00
68 T	m/p-Xylenes	0.920	0.963	-4.7	89	0.00
69 T	o-Xylene	0.907	0.956	-5.4	90	0.00
70 T	Styrene	1.490	1.630	-9.4	90	0.00
71 P	Bromoform	0.339	0.372	-9.7	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.903	3.936	-0.8	90	0.00
74 T	N-amyl acetate	1.897	2.145	-13.1	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.429	1.514	-5.9	97	0.00
76 T	1,2,3-Trichloropropane	1.318	1.367	-3.7	94	0.00
77 T	Bromobenzene	0.916	0.895	2.3	87	0.00
78 T	n-propylbenzene	4.659	4.777	-2.5	89	0.00
79 T	2-Chlorotoluene	2.942	2.866	2.6	90	0.00
80 T	1,3,5-Trimethylbenzene	3.268	3.328	-1.8	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.439	0.389	11.4	74	0.00
82 T	4-Chlorotoluene	3.214	3.252	-1.2	90	0.00
83 T	tert-Butylbenzene	3.029	3.213	-6.1	93	0.00
84 T	1,2,4-Trimethylbenzene	3.261	3.396	-4.1	89	0.00
85 T	sec-Butylbenzene	4.146	4.316	-4.1	90	0.00
86 T	p-Isopropyltoluene	3.317	3.438	-3.6	90	0.00
87 T	1,3-Dichlorobenzene	1.782	1.758	1.3	90	0.00
88 T	1,4-Dichlorobenzene	1.814	1.757	3.1	89	0.00
89 T	n-Butylbenzene	3.088	3.419	-10.7	94	0.00
90 T	Hexachloroethane	0.571	0.568	0.5	88	0.00
91 T	1,2-Dichlorobenzene	1.749	1.842	-5.3	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.343	0.398	-16.0	106	0.00
93 T	1,2,4-Trichlorobenzene	1.014	1.115	-10.0	98	0.00
94 T	Hexachlorobutadiene	0.407	0.421	-3.4	94	0.00
95 T	Naphthalene	3.962	4.617	-16.5	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080122\
Data File : VX030338.D
Acq On : 01 Aug 2022 21:22
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 02 01:58:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.040	1.142	-9.8	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6