

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080123\
 Data File : VX036743.D
 Acq On : 01 Aug 2023 13:01
 Operator : JC/MD
 Sample : 03808-05DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1712DL

Quant Time: Aug 02 06:24:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	132164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	241575	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100704	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90105	46.076	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.160%		
35) Dibromofluoromethane	5.391	113	77119	46.445	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.880%		
50) Toluene-d8	8.647	98	288687	49.027	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.060%		
62) 4-Bromofluorobenzene	11.079	95	108575	48.873	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.740%		
Target Compounds						
16) Acetone	2.380	43	2374	2.926	ug/l	92
20) Methylene Chloride	2.788	84	1180	0.654	ug/l	94
25) 2-Butanone	4.568	43	1849	1.404	ug/l	98
52) Toluene	8.720	92	359777	79.741	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080123\
Data File : VX036743.D
Acq On : 01 Aug 2023 13:01
Operator : JC/MD
Sample : 03808-05DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1712DL

Quant Time: Aug 02 06:24:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

