

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080123\
 Data File : VX036745.D
 Acq On : 01 Aug 2023 13:49
 Operator : JC/MD
 Sample : 03808-07DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1714DL

Quant Time: Aug 02 06:24:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138652	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246971	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92837	45.252	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.500%
35) Dibromofluoromethane	5.385	113	79018	46.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.100%
50) Toluene-d8	8.647	98	293709	48.790	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	102813	45.268	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.540%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1708	2.006	ug/l	97
20) Methylene Chloride	2.794	84	1118	0.591	ug/l #	71
25) 2-Butanone	4.574	43	2745	1.987	ug/l	97
52) Toluene	8.720	92	277684	60.202	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	1951	0.305	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080123\
Data File : VX036745.D
Acq On : 01 Aug 2023 13:49
Operator : JC/MD
Sample : 03808-07DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1714DL

Quant Time: Aug 02 06:24:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

