

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080123\
 Data File : VX036750.D
 Acq On : 01 Aug 2023 15:50
 Operator : JC/MD
 Sample : 03856-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-04

Quant Time: Aug 02 06:25:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130595	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	234744	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87602	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88039	45.561	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.120%
35) Dibromofluoromethane	5.391	113	74564	46.213	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.420%
50) Toluene-d8	8.647	98	277555	48.508	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.020%
62) 4-Bromofluorobenzene	11.079	95	98428	45.595	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.180%
Target Compounds						
20) Methylene Chloride	2.788	84	782	0.439	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080123\
Data File : VX036750.D
Acq On : 01 Aug 2023 15:50
Operator : JC/MD
Sample : 03856-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-04

Quant Time: Aug 02 06:25:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

