

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080123\
 Data File : VX036758.D
 Acq On : 01 Aug 2023 19:03
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 02 06:27:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	224939	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84258	46.011	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.020%
35) Dibromofluoromethane	5.385	113	71426	46.197	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.400%
50) Toluene-d8	8.647	98	269530	49.159	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.320%
62) 4-Bromofluorobenzene	11.079	95	100882	48.769	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.540%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1126	0.666	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080123\
Data File : VX036758.D
Acq On : 01 Aug 2023 19:03
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 02 06:27:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

