

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042796.D
 Acq On : 01 Aug 2024 15:25
 Operator : JC/MD
 Sample : P3387-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-9-Full Waste Class

Quant Time: Aug 01 16:46:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	58712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	125443	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46579	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61785	65.502	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	131.000%#
35) Dibromofluoromethane	5.391	113	44014	50.791	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.580%
50) Toluene-d8	8.647	98	165416	52.589	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.180%
62) 4-Bromofluorobenzene	11.079	95	59164	51.470	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.940%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	39468	115.231	ug/l	93
18) Methyl Acetate	2.709	43	1889	2.197	ug/l	94
20) Methylene Chloride	2.782	84	4965	5.581	ug/l	88
25) 2-Butanone	4.593	43	4533	9.008	ug/l	93
43) Isopropyl Acetate	6.342	43	72415	40.358	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042796.D
Acq On : 01 Aug 2024 15:25
Operator : JC/MD
Sample : P3387-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-9-Full Waste Class

Quant Time: Aug 01 16:46:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

