

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042799.D
 Acq On : 01 Aug 2024 16:56
 Operator : JC/MD
 Sample : P3387-03RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-9-Full Waste ClassRE

Quant Time: Aug 01 18:12:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 08/03/2024
 Supervised By :Semsettin Yesilyurt 08/03/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	138920	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117941	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43989	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51126	47.486	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.980%		
35) Dibromofluoromethane	5.385	113	40140	41.827	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	83.660%		
50) Toluene-d8	8.647	98	147553	42.359	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	84.720%#		
62) 4-Bromofluorobenzene	11.079	95	58378	45.859	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.720%		
Target Compounds					Qvalue	
16) Acetone	2.392	43	31064	79.458	ug/l	91
18) Methyl Acetate	2.709	43	3528	3.596	ug/l #	87
20) Methylene Chloride	2.782	84	4307	4.044	ug/l #	79
25) 2-Butanone	4.586	43	3728m	6.490	ug/l	
43) Isopropyl Acetate	6.342	43	63163	31.786	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042799.D
Acq On : 01 Aug 2024 16:56
Operator : JC/MD
Sample : P3387-03RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-9-Full Waste ClassRE

Quant Time: Aug 01 18:12:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/03/2024
Supervised By :Semsettin Yesilyurt 08/03/2024

