

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042801.D
 Acq On : 01 Aug 2024 17:43
 Operator : JC/MD
 Sample : P3396-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-CF02-01

Quant Time: Aug 02 03:15:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	61581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	126042	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47896	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	55631	56.230	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 112.460%			
35) Dibromofluoromethane	5.391	113	44183	50.744	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.480%			
50) Toluene-d8	8.647	98	154509	48.888	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 97.780%			
62) 4-Bromofluorobenzene	11.079	95	55636	48.171	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.340%			
Target Compounds						
16) Acetone	2.392	43	4700	13.083	ug/l	94
20) Methylene Chloride	2.788	84	15154	17.817	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042801.D
Acq On : 01 Aug 2024 17:43
Operator : JC/MD
Sample : P3396-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-CF02-01

Quant Time: Aug 02 03:15:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

