

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042803.D
 Acq On : 01 Aug 2024 18:30
 Operator : JC/MD
 Sample : P3396-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-CF04-01

Quant Time: Aug 02 03:16:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	59963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	125816	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	57249	59.426	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	118.860%#
35) Dibromofluoromethane	5.385	113	43861	50.464	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.920%
50) Toluene-d8	8.647	98	148554	47.088	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.180%
62) 4-Bromofluorobenzene	11.079	95	55234	47.909	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.820%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	3953	11.300	ug/l	91
20) Methylene Chloride	2.788	84	14993	18.117	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042803.D
Acq On : 01 Aug 2024 18:30
Operator : JC/MD
Sample : P3396-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-CF04-01

Quant Time: Aug 02 03:16:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

