

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042805.D
 Acq On : 01 Aug 2024 19:17
 Operator : JC/MD
 Sample : P3396-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-CF01-01

Quant Time: Aug 02 03:17:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	53818	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	112965	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51105	59.106	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 118.220%#	
35) Dibromofluoromethane	5.385	113	40291	51.630	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.260%	
50) Toluene-d8	8.647	98	128938	45.520	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 91.040%#	
62) 4-Bromofluorobenzene	11.079	95	49219	47.548	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.100%	
Target Compounds						Qvalue
16) Acetone	2.392	43	4351	13.858	ug/l	94
20) Methylene Chloride	2.788	84	11921	15.955	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042805.D
Acq On : 01 Aug 2024 19:17
Operator : JC/MD
Sample : P3396-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-CF01-01

Quant Time: Aug 02 03:17:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

