

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042806.D
 Acq On : 01 Aug 2024 19:40
 Operator : JC/MD
 Sample : P3396-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-CF05-01

Quant Time: Aug 02 03:17:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	57671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	117288	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51898	56.013	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 112.020%	
35) Dibromofluoromethane	5.391	113	39197	48.377	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.760%	
50) Toluene-d8	8.647	98	147333	50.097	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.200%	
62) 4-Bromofluorobenzene	11.079	95	53974	50.220	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.440%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3124	9.285	ug/l #	76
20) Methylene Chloride	2.788	84	20431	26.013	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042806.D
Acq On : 01 Aug 2024 19:40
Operator : JC/MD
Sample : P3396-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-CF05-01

Quant Time: Aug 02 03:17:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

