

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042812.D
 Acq On : 01 Aug 2024 23:35
 Operator : JC/MD
 Sample : VX0801WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0801WBL02

Quant Time: Aug 02 03:22:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	59138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	119066	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	43889	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	52935	55.715	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.420%
35) Dibromofluoromethane	5.391	113	40789	49.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.180%
50) Toluene-d8	8.646	98	144705	48.469	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.940%
62) 4-Bromofluorobenzene	11.079	95	52328	47.961	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042812.D
Acq On : 01 Aug 2024 23:35
Operator : JC/MD
Sample : VX0801WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0801WBL02

Quant Time: Aug 02 03:22:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

