

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042817.D
 Acq On : 02 Aug 2024 01:55
 Operator : JC/MD
 Sample : P3374-25
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-02-072024

Quant Time: Aug 02 05:40:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	122835	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116115	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45400	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	52648	51.864	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.720%
35) Dibromofluoromethane	5.385	113	42229	49.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.540%
50) Toluene-d8	8.647	98	152527	49.521	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.040%
62) 4-Bromofluorobenzene	11.079	95	53436	47.474	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.940%
Target Compounds						
16) Acetone	2.392	43	2665	7.230	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042817.D
Acq On : 02 Aug 2024 01:55
Operator : JC/MD
Sample : P3374-25
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-02-072024

Quant Time: Aug 02 05:40:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

