

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
 Data File : VX042818.D
 Acq On : 02 Aug 2024 02:18
 Operator : JC/MD
 Sample : P3430-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-02-073124

Quant Time: Aug 02 05:41:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 01 16:40:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	145279	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	139591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58105	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	58330	48.518	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.040%
35) Dibromofluoromethane	5.385	113	48096	47.923	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.840%
50) Toluene-d8	8.647	98	177509	48.729	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.460%
62) 4-Bromofluorobenzene	11.079	95	65705	49.356	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.720%
Target Compounds						
16) Acetone	2.392	43	1355	3.104	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080124\
Data File : VX042818.D
Acq On : 02 Aug 2024 02:18
Operator : JC/MD
Sample : P3430-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-02-073124

Quant Time: Aug 02 05:41:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X080124W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 01 16:40:12 2024
Response via : Initial Calibration

